

Application News



Probe disinfection

Which disinfection agents can be used for the probes

With every probe a paper with all proven disinfection agents is delivered in the transportation box. This paper gives an overview on all agents proven at this time. Since disinfection agents may change from time to time here a link to get an updated version of this.

Click on the link:

http://www.gehealthcare.com/usen/ultrasound/products/probe_care.html

Here is some general information about probe care.

This site provides updated probe care cards for all GE ultrasound systems (Vivid, Logiq, Voluson).

Select Vivid Transducer care, to get all Vivid related probes.

Be aware this is an overview on all probes that can be used on different vivid systems. Not all the probes are running on all the systems.

Select the probe that is needed.

A new window will open where a pdf document is shown.

In some documents an overview of several probes is shown.



In order to store this – click on the store icon –



Select the place of storage – change name if needed and press Save.